

CERTIFICATE OF CALIBRATION

sales | hire | calibration | repair of new / secondhand weighing equipment

Certificate Number: **17699**

Customer Name:	Scobie Junor
Site Address:	
	1 Singer Rod
	East Kilbride
	Scotland
	G75 0XS
Contact Name:	
Telephone No:	

EQUIPMENT DETAILS:	
Equipment ID.	16
Equipment	Compounding
Make.	Uni Argeo
Model.	590EGTC3GD
Serial No.	83517264
Capacity.	1500g/30000g
Capacity	0.01g/0.1g
Tolerance	+/- 0.03g/0/3g

LINEARITY:		
Applied Weight	Reading As Found	Reading As Left (Definitive)
Zero	0.00g	0.00g
100g	100.00g	100.00g
500g	499.99g	500.00g
1500g	1499.94g	1500.00g
2000g	2000.0g	2000.0g
8000g	8000.0g	8000.0g
16000g	15999.7g	16000.0g
30000g	29999.1g	30000.0g

CALIBRATION DETAILS:	
Test Weight Numbers.	W3705 W3706 SWS 9-31 YB 8-15
Test Weight Certificate Numbers.	39861 39860 3560

ECCENTRIC LOADING:			
Applied Weight		Reading As Found	Reading As Left (Definitive)
Corner 1.	8000g	8000.0g	8000.0g
Corner 2.	8000g	8000.0g	8000.0g
Corner 3.	8000g	8000.0g	8000.0g
Corner 4.	8000g	8000.1g	8000.0g

REPEATABILITY (x3) TEST POINTS:		
Applied Weight	Reading As Found	Reading As Left (Definitive)
8000g	8000.0g	8000.0g
8000g	8000.0g	8000.0g
8000g	8000.1g	8000.0g
Comments: None		

has been tested and calibrated in accordance with your stated Order requirements. We further certify that the above equipment has been calibrated using equipment and Standards that are traceable to National Standards maintained by the national weights and measures laboratory.

CERTIFICATE ISSUED BY:		CUSTOMERS SIGNATURE IF APPLICABLE	
Name:	Darren Heald	Name:	
Signature:	<i>Darren Heald</i>	Signature:	
Date:	28th September 2021	Date:	
		CALIBRATION DATE:	
		28th September 2021	
		CALIBRATION NEXT DUE:	
		September 2022	